

§ 486.102

(d) *Standard—conformity with other Federal, State, and local laws and regulations.* The supplier of portable X-ray services agrees to render such services in conformity with Federal, State, and local laws relating to safety standards.

[34 FR 388, Jan. 10, 1969. Redesignated at 42 FR 52826, Sept. 30, 1977. Further redesignated and amended at 60 FR 2326, Jan. 9, 1995; 60 FR 45086, Aug. 30, 1995]

§ 486.102 Condition for coverage: Supervision by a qualified physician.

Portable X-ray services are provided under the supervision of a qualified physician.

(a) *Standard—physician supervision.* The performance of the roentgenologic procedures is subject to the supervision of a physician who meets the requirements of paragraph (b) of this section and one of the following requirements is met:

(1) The supervising physician owns the equipment and it is operated only by his employees, or

(2) The supervising physician certifies annually that he periodically checks the procedural manuals and observes the operators' performance, that he has verified that equipment and personnel meet applicable Federal, State, and local licensure and registration requirements and that safe operating procedures are used.

(b) *Standard—qualifications of the physician supervisor.* Portable X-ray services are provided under the supervision of a licensed doctor of medicine or licensed doctor of osteopathy who is qualified by advanced training and experience in the use of X-rays for diagnostic purposes, i.e., he (1) is certified in radiology by the American Board of Radiology or by the American Osteopathic Board of Radiology or possesses qualifications which are equivalent to those required for such certification, or (2) is certified or meets the requirements for certification in a medical specialty in which he has become qualified by experience and training in the use of X-rays for diagnostic purposes, or (3) specializes in radiology

42 CFR Ch. IV (10–1–14 Edition)

and is recognized by the medical community as a specialist in radiology.

[34 FR 388, Jan. 10, 1969. Redesignated at 42 FR 52826, Sept. 30, 1977. Further redesignated and amended at 60 FR 2326, Jan. 9, 1995; 60 FR 45086, Aug. 30, 1995]

§ 486.104 Condition for coverage: Qualifications, orientation and health of technical personnel.

Portable X-ray services are provided by qualified technologists.

(a) *Standard—qualifications of technologists.* All operators of the portable X-ray equipment meet the requirements of paragraphs (a)(1), (2), (3), or (4) of this section:

(1) Successful completion of a program of formal training in X-ray technology in a school approved by the Joint Review Committee on Education in Radiologic Technology (JRCERT), or have earned a bachelor's or associate degree in radiologic technology from an accredited college or university.

(2) For those whose training was completed prior to July 1, 1966, but on or after July 1, 1960: Successful completion of 24 full months of training and/or experience under the direct supervision of a physician who is certified in radiology by the American College of Radiology or who possesses qualifications which are equivalent to those required for such certification, and at least 12 full months of pertinent portable X-ray equipment operation experience in the 5 years prior to January 1, 1968.

(3) For those whose training was completed prior to July 1, 1960: Successful completion of 24 full months of training and/or experience of which at least 12 full months were under the direct supervision of a physician who is certified in radiology by the American College of Radiology or who possesses qualifications which are equivalent to those required for such certification, and at least 12 full months of pertinent portable X-ray equipment operation experience in the 5 years prior to January 1, 1968.

(4) For those whose training was completed prior to January 1, 1993, successful completion of a program of formal training in X-ray technology in a

school approved by the Council on Education of the American Medical Association, or by the American Osteopathic Association is acceptable.

(b) *Standard—personnel orientation.* The supplier of portable X-ray services has an orientation program for personnel, based on a procedural manual which is: Available to all members of the staff, incorporates relevant portions of professionally recognized documents, and includes instruction in all of the following:

(1) Precautions to be followed to protect the patient from unnecessary exposure to radiation;

(2) Precautions to be followed to protect an individual supporting the patient during X-ray procedures from unnecessary exposure to radiation;

(3) Precautions to be followed to protect other individuals in the surrounding environment from exposure to radiation;

(4) Precautions to be followed to protect the operator of portable X-ray equipment from unnecessary exposure to radiation;

(5) Considerations in determining the area which will receive the primary beam;

(6) Determination of the time interval at which to check personnel radiation monitors;

(7) Use of the personnel radiation monitor in providing an additional check on safety of equipment;

(8) Proper use and maintenance of equipment;

(9) Proper maintenance of records;

(10) Technical problems which may arise and methods of solution;

(11) Protection against electrical hazards;

(12) Hazards of excessive exposure to radiation.

(c) *Standard: Employee records.* Records are maintained and include evidence that—

(1) Each employee is qualified for his or her position by means of training and experience; and

(2) Employees receive adequate health supervision.

[34 FR 388, Jan. 10, 1969. Redesignated at 42 FR 52826, Sept. 30, 1977, and amended at 53 FR 12015, Apr. 12, 1988; 60 FR 45086, Aug. 30, 1995; 73 FR 69942, Nov. 19, 2008]

§ 486.106 Condition for coverage: Referral for service and preservation of records.

All portable X-ray services performed for Medicare beneficiaries are ordered by a physician or a nonphysician practitioner as provided in § 410.32(a) of this chapter or by a nonphysician practitioner as provided in § 410.32(a)(2) and records are properly preserved.

(a) *Standard—referral by a physician or nonphysician practitioners.* Portable X-ray examinations are performed only on the order of a physician licensed to practice in the State or by a nonphysician practitioner acting within the scope of State law. Such nonphysician practitioners may be treated the same as physicians treating beneficiaries for the purpose of this paragraph. The supplier's records show that:

(1) The portable X-ray test was ordered by a licensed physician or a nonphysician practitioner acting within the State scope of law; and

(2) Such physician or nonphysician practitioner's written, signed order specifies the reason a portable X-ray test is required, the area of the body to be exposed, the number of radiographs to be obtained, and the views needed; it also includes a statement concerning the condition of the patient which indicates why portable X-ray services are necessary.

(b) *Standard—records of examinations performed.* The supplier makes for each patient a record of the date of the portable X-ray examination, the name of the patient, a description of the procedures ordered and performed, the referring physician or nonphysician practitioner, the operator(s) of the portable X-ray equipment who performed the examination, the physician to whom the radiograph was sent, and the date it was sent.

(c) *Standard—preservation of records.* Such reports are maintained for a period of at least 2 years, or for the period of time required by State law for such records (as distinguished from requirements as to the radiograph itself), whichever is longer.

[34 FR 388, Jan. 10, 1969. Redesignated at 42 FR 52826, Sept. 30, 1977. Further redesignated and amended at 60 FR 2326, Jan. 9, 1995; 60 FR 45086, Aug. 30, 1995; 77 FR 69372, Nov. 16, 2012]